

LBS COLLEGE OF ENGINEERING, KASARAGOD

QUOTATION NOTICE

No. D-1156/2021 BSK-5

Date: 17.08.2022

Sealed competitive quotations are invited for the purchase of **Laboratory Meters for Electrical Machines Lab** in this institution specified in the schedule attached. Super scribing quotations for the supply of materials against quotation **No.D-1156/2021 BSK-5** due on 13/09/2022. The rates quoted should be for the delivery of the items at the college. Quotations received after the time fixed on the due date is liable to be rejected. Maximum period required for the delivery of items should also be mentioned.

The prices quoted should be inclusive of all taxes, duties etc. The quotationer shall also quote the percentage of rebate/discount if any offered by them against the items.

Special conditions if any or printed on quotation notice will not be applicable to the contract unless they are expressly accepted in writing by the undersigned.

Due date and time for Receipt of Quotation : 13/09/2022 at 2 pm
Date and time of opening the Quotation : 13/09/2022 at 3 pm
The Designation and address of the Officer : **The Principal**
To whom the Quotation is to be addressed : LBS College of Engineering,
Muliyar (P.O), Kasaragod-
671 542




PRINCIPAL

Encl: Specification of item
Copy to:
1) College website
2) Notice Board

Specification of Items			
Sl No.	Equipment/Item	Description	Quantity
1.	Voltmeter	(0-150/300/600V) MI, Display- Dial type, Anti-parallax mirror and knife edged pointer, ABS Moulded plastic cabinet. Accuracy – Class 1, Warranty – 1 Year, Table top type.	5
2.		(0-300V) MC, Display- Dial type, Anti-parallax mirror and knife edged pointer, ABS Moulded plastic cabinet. . Accuracy – Class 1, Warranty – 1 Year, Table top type.	4
3.	Ammeter	(0-2.5/5/10A) MI, Display- Dial type, Anti-parallax mirror and knife edged pointer, ABS Moulded plastic cabinet. . Accuracy – Class 1, Warranty – 1 Year, Table top type.	5
4.		(0-10/20A) MC, Display- Dial type, Anti-parallax mirror and knife edged pointer, ABS Moulded plastic cabinet. . Accuracy – Class 1, Warranty – 1 Year, Table top type.	4
5.		(0-1/2A) MC, Display- Dial type, Anti-parallax mirror and knife edged pointer, ABS Moulded plastic cabinet. . Accuracy – Class 1, Warranty – 1 Year, Table top type.	4
6.	Wattmeter	(0-300/600V, 2.5/5A LPF), Accuracy - +/- 1.5%, ISS 1248/83, Display- Dial type, Anti-parallax mirror and knife edged pointer, Enclosed in wooden cabinet, Warranty – 1 Year, Table top type.	6
7.		(0-300/600V, 5/10A UPF), Accuracy - +/- 1.5%, ISS 1248/83, Display- Dial type, Anti-parallax mirror and knife edged pointer, Enclosed in wooden cabinet, Warranty – 1 Year, Table top type, Table top type.	6
8.		(0-75/150/300V, 2.5/5A, UPF) , Accuracy - +/- 1.5%, ISS 1248/83, Display- Dial type, Anti-parallax mirror and knife edged pointer, Enclosed in wooden cabinet, Warranty – 1 Year, Table top type.	2


PRINCIPAL
PRINCIPAL
L.B.S COLLEGE OF ENGINEERING
KASARAGOD - 671 542