

LBS COLLEGE OF ENGINEERING, KASARAGOD

QUOTATION NOTICE

No. D-1156/2021 BSK-7

Date: 22.08.2022

Sealed competitive quotations are invited for the purchase of **1 Number of Low Current Clamp Meter** in this institution specified in the schedule attached. Super scribing quotations for the supply of materials against quotation **No.D-1156/2021 BSK-7** due on 15/09/2022. The rates quoted should be for the delivery of the items at the college. Quotations received after the time fixed on the due date is liable to be rejected. Maximum period required for the delivery of items should also be mentioned.

The prices quoted should be inclusive of all taxes, duties etc. The quotationer shall also quote the percentage of rebate/discount if any offered by them against the items.

Special conditions if any or printed on quotation notice will not be applicable to the contract unless they are explicitly accepted in writing by the undersigned.

Due date and time for Receipt of Quotation	:	15/09/2022 at 2 pm
Date and time of opening the Quotation	:	15/09/2022 at 3 pm
The Designation and address of the Officer	:	The Principal
Date up to which the rates are to remain firm for acceptance	:	6 Months
To whom the Quotation is to be addressed	:	LBS College of Engineering, Muliyar (P.O), Kasaragod- 671 542

Encl: Specification of item

Copy to:

- 1) College website
- 2) Notice Board.




PRINCIPAL
PRINCIPAL
L.B.S COLLEGE OF ENGINEERING
KASARAGOD - 671 542

Technical Specifications for Low Current Clamp Meter (1 No)

Display: 6000 counts (Min)
TRMS Leakage current and load current
Range: 8.000 mA to 100.0 A
Basic accuracy: $\pm (1.0 \% + 3d)$

AC Current Range	Resolution	Accuracy
8 mA	0.001 mA	
600 mA	0.1 mA	0 ~ 10 A $\pm(1 \% \text{ reading} + 8 \text{ digits})$ 10 ~ 50 A
6 A	0.001 A	$\pm(2 \% \text{ reading} + 10 \text{ digits})$ 50 ~ 100 A
60 A	0.01 A	
100 A	0.1 A	$\pm(10 \% \text{ reading} + 10 \text{ digits})$

- Operating frequency range: - 50 - 400 Hz
- Data, auto, peak hold: User Selectable
- Low pass filter: 50 / 60 Hz
- Auto, data & peak hold
- Jaw size: 40 mm (Min)
- Overload: "OL"
- Low battery:
- Auto power off: 30 minutes of inactivity
- Operating temperature: 0 °C - 40 °C
- Power Supply: Internal Battery
- Battery life: 60 hours
- Safety: EN61010-1 CAT III 300 V
- EMC: EN61326-1
- Warranty - 1 Year Minimum

Handwritten signature

PRINCIPAL

PRINCIPAL
L.B.S COLLEGE OF ENGINEERING
KASARAGOD - 671 542

